

Message 001

Communication from the Commission - TRIS/(2026) 2596

Directive (EU) 2015/1535

Notification: 2026/0507/RO

Notification of a draft text from a Member State

Notification – Notification – Notifizierung – Нотификация – Oznámení – Notifikation – Γνωστοποίηση – Notificación – Teavitamine – Ilmoitus – Obavijest – Bejelentés – Notifica – Pranešimas – Paziņojums – Notifika – Kennisgeving – Zawiadomienie – Notificação – Notificare – Oznámenie – Obvestilo – Anmälan – Fógra a thabhairt

Does not open the delays - N'ouvre pas de délai - Kein Fristbeginn - Не се предвижда период на прекъсване - Nezahtuje prodlení - Fristerne indledes ikke - Καμμία έναρξη προθεσμίας - No abre el plazo - Viivituste perioodi ei avata - Määräaika ei ala tästä - Ne otvara razdoblje kašnjenja - Nem nyitja meg a késésekét - Non fa decorrere la mora - Atidėjimai nepradedami - Atlikšanas laikposms nesākas - Ma jiftaħx il-perijodi ta' dewmien - Geen termijnbegin - Nie otwiera opóźnień - Não inicia o prazo - Nu deschide perioadele de stagnare - Nezačína oneskorenia - Ne uvaja zamud - Inleder ingen frist - Ní osclaíonn sé na moilleanna

MSG: 20262596.EN

1. MSG 001 IND 2026 0507 RO EN 16-09-2026 RO NOTIF

2. Romania

3A. Ministerul Economiei, Digitalizării, Antreprenoriatului și Turismului
Calea Victoriei nr.152, Sector 1, Bucuresti, cod 010096;
Telefon 004 0372492634
E-mail: reglementari_tehnice@economie.gov.ro

3B. Biroul Român de Metrologie Legală
Tel: +40.21.332.09.54
Adresă: os. Vitan Bârzeti 11, Sector 4, cod 042122, Bucureti, România
E-mail: office@brml.ro

4. 2026/0507/RO - I10 - Metrology

5. Draft ORDER amending and supplementing the annex to Order No 77/2022 of the Director-General of the Romanian Bureau of Legal Metrology (BRML) approving the Official list of measuring instruments subject to legal metrological control LO –2022

6. The draft concerns the inclusion of ultrasonic gas meters with a maximum flow rate of up to 40 m³/h,

7.

8. The draft concerns:

- the inclusion of ultrasonic gas meters with a maximum flow rate of up to 40 m³/h, including in the category of measuring instruments subject to legal metrological control in Romania; it establishes the procedures for legal metrological control (periodic metrological inspections as well as metrological inspections following repair or modification) and the periodicity of the legal metrological control conducted through the periodic metrological inspection
- extension of the maximum flow rate for turbine gas meters from 2 500 m³/h up to 6 500 m³/h

9. Due to advances in the technical capability available at the National Institute of Metrology to carry out metrological inspections of ultrasonic gas meters with a maximum flow rate of up to 40 m³/h, inclusive, as well as at laboratories authorised by the Romanian Bureau of Legal Metrology to carry out metrological inspections of turbine gas meters with a maximum flow rate of up to 6 500 m³/h, it is necessary to include them in the Official list of measuring instruments subject to legal metrological control LO – 2022, approved by Order No 77/2022 of the Director-General of the Romanian Bureau of Legal Metrology, published in the Official Gazette of Romania, Part I, No 332/05.04.2022, amending Order No 204/2024 of the Director-General of the Romanian Bureau of Legal Metrology, so that they may, once they have been made available on the market and put into service, be subject to legal metrological control conducted, for example, through periodic metrological inspections or following repair, in order to ensure the accuracy of the measurements made in commercial transactions in which they are used.

We would like to point out that both ultrasonic gas meters and turbine gas meters are subject, prior to being placed on the market, to an assessment of conformity with the provisions of Government Decision No 711/2015 establishing the conditions for placing measuring instruments on the market, as subsequently amended and supplemented, which decision transposes the provisions of Directive No 2014/32/EU, as subsequently amended and supplemented, into national law.

9a. The proposal to include ultrasonic gas meters with a maximum flow rate up to 40 m³/h, as well as turbine gas meters with a maximum flow rate between 2 500 m³/h and 6 500 m³/h, inclusive, in the category of measuring instruments subject to legal metrological control in Romania, establishes the procedures for legal metrological control and is adequate to ensure the accuracy of the measurements made in commercial transactions.

9b. The proposed amendment is necessary in order to update the Official List of measuring instruments subject to legal metrological control LO – 2022 and does not impose additional restrictions on the internal market of the European Union; nor does it affect cross-border trade or services.

9c. The proposed amendment does not constitute a disproportionate burden in relation to the intended objective and does not impose any restrictions beyond those laid down in Directive No 2014/32/EU, as subsequently amended and supplemented, and international regulations; it also fulfils the intended objective.

10. Reference(s) to basic text(s): The basic texts were forwarded with an earlier notification:
2024/0389/RO
2021/0841/RO

11. No

12.

13. No

14. No

15. No

16.

TBT aspects: No

SPS aspects: No

European Commission

Contact point Directive (EU) 2015/1535

email: grow-dir2015-1535-central@ec.europa.eu